

## **HI-VAC ADAPTER**

Hi-Vac adapter for use with the OTA100 Osseous Collectors.

If you have a Hi-Vac suction you will need to use OTA's Hi-Vac Adapter with your OTA100 Osseous Collector.

The Hi-Vac Adapter is provided non-sterile and should be cleaned and sterilized using standard cleaning and autoclaving procedures prior to use.

You may autoclave and re-use the Hi-Vac adapter. We recommend replacement after about 20 uses.